

Digital Linear Gauge Sensors

A selection of eight models suitable for installation in production lines with environments subject to dust, water splashes or/and oil splashes.

Features

- Conforming to environmental protection class IP64 or IP66G.
- Various optional parts.
- Compatible with DG-4000 series.
- Compact and small body with high accuracy.
- High cost performance.

Structure Conforms to the IP64/66G Protection Class

The IP64/66G International Protection number code indicates the protection class with respect to the penetration of dust, water, and oil

IP6X indicates that the enclosure is dust-tight, with no ingress of dust

IPX4 indicates that the enclosure is protected against drops of water and splashing water. Water splashed against the enclosure from any direction shall

IPX6 indicates that the enclosure is protected against heavy seas, or powerful jets of water. With no ingress of water jets from any direction.

IPXXG indicates that the enclosure is protected against oil drops and oil splashes. Oil splashed against the enclosure from any direction shall have no

■ Compatible digital counters

DG-4120 BCD output

DG-4140 Comparator output



(BCD, Color backlight)



DG-4190 Comparator output (Analog, Color backlight)



DG-4240 Comparator output (Upper & Lower)



DG-4280 Comparator output (4-preset)




Error = d' - d

■ A Brief Explanation of Terms

Display Accuracy

This is the measurement error inherent in a linear gauge sensor. The error (the amount of difference from the actual value) is measured at each specified measurement value, and, when the overall length of the spindle movement is considered, the sum of the absolute values of the maximum error in the positive direction and of the maximum error in the negative direction becomes the display accuracy of that gauge sensor (see Fig. 1).

Measurement of the accuracy is performed by making comparisons with a reference displacement meter. The difference between the reading value at the time that the lowest order digit of the target sensor changed and the value of the reference displacement meter is taken as the error. This is the reason why the display accuracy of a sensor with a resolution of 10µm is lower than that of a sensor with a resolution of 3µm. (see Fig. 2).



The force used to hold down the workpiece is called the measurement force. Since our gauge sensors feature an internal spring-return mechanism for the spindle, the measurement force is the force measured at the maximum extension limit. The force unit is expressed as N (Newton). The measurement force can be changed by replacing the spring. Please specify your requirements when placing your order. Please note that depending on the modification, there may be times when the attachment does not return completely if it is facing upwards or sideways. Likewise, you will need to consult us if modification to a fixed value (see*1 below) is required.

*1: If, for example, modification to a fixed value for a linear gauge sensor with a measurement range of 13mm was made, the measurement force for the amount moved within the 0 to 13mm range becomes a fixed value (variation of ± 10 gf).

GS-4700/4800/6700/6800 series

Output signal circuit

GS-4713/4730/4813/4830/6713/6730/6813/6830


Connector: R03-PB6M (product of TAJIMI ELECTRONICS CO., LTD.) (Line color) +5V ^ SIG 1 (Yellow) В SIG 2 (White) +5V IN (Orange) C D No connection Е COM (Grav) COM (Pink) 22Ω

Signal connector R03-PB6M

F A	Pin No.	Signal name	Line color
E B	Α	SIG1	Yellow
	В	SIG2	White
	С	+5V IN	Orange
D C	D	NC	Not connected
Pin assignment	E	COM	Gray
	F	COM	Pink

* At Ono Sokki, we recommend the use of a linear gauge sensor together with a digital gauge counter. If you plan to use a linear gauge sensor without an Ono Sokki digital gauge counter, please refer to the specifications in the user's manual for the sensor and/or other materials to design your o

Outer Dimensions




■ Specification

Model name	GS-4713	GS-4730	GS-4813	GS-4830	GS-6713	GS-6730	GS-6813	GS-6830
Measurement range	13mm	30mm	13mm	30mm	13mm	30mm	13mm	30mm
Resolution	10µm		1µm		10μm		1µm	
Accuracy (at +20°C)	3μm		2µm	3µm	3µm	3µm	2µm	3µm
Maximum spindle velocity *1	1 (4) m/s		0.3 (1.2) m/s		1 (4) m/s		0.3 (1.2) m/s	
Measurement force *2 (downward)	2.0N or less	2.4N or less	2.0N or less	2.4N or less	1.7N or less	2.0N or less	1.7N or less	2.0N or less
Number of sliding times (proven in our endurance test)	15 million or more			5 million or more				
Protection class (excluding connector section)	IP66G			IP64				
Stem diameter	ø15- ⁴⁰ ₋₀₀₀₉ mm							
Power requirement	4.5 to 5.5VDC							
Power consumption (when 5VDC)	120mA or less							
Signal output (when 5VDC)	Two-phase square wave, Phase difference: 90°± 20°, Output voltage Hi: 4.5V or more Lo: 0.4V or less							
Output impedance	Approx. 22Ω							
Vibration resistance (when the power is off)*3	196m/s² in each of three axial directions (for 75 minutes each) 10 cycles of 10 to 150Hz sweep 10 cycles of 10 to 150Hz sweep				inutes each)			
Shock resistance (when the power is off)*3	1960m/s² in each direction for each of three axes, three times for each of ±X, Y and Z directions, sine half-wave, pulse duration 6ms each of ±X, Y and Z directions, sine half-wave, pulse duration							
Operating temperature range	0 to +40°C							
Storage temperature range	-10 to +55°C							
Cable length	Approx. 5m Approx. 1.9 m							
Weight (including cable and connector)	Approx. 325g	Approx. 385g	Approx. 325g	Approx. 385g	Approx. 250g	Approx.310g	Approx. 250g	Approx. 310g

- *1: When used with Ono Sokki's digital gauge counter. The values within parentheses () are the maximum spindle velocity used with the DG-4140/4190/5100.
- *2: An attachment may not return completely if it is facing upwards.
 *3: Vibration / shock resistance values described in above are not guaranteed during measurement operation.
- Disconnected or modified signal cable is not applicable to CE marking.

Contact tip




^{*1} When affixing a flat gauge head such as the AA-0220/0240/0250/921 to a gauge with a measurement resolution of 0.05µm/1µm, adjustment of the degree of parallelization to match that of the surface of the measurement stand is required. In this case, the gauge head and stand must be purchased as a pair (additional cost required).


Extension spindle

Model name	AA-844	AA-845	
	Material: SUS303	Material: SUS303	
Outer dimensions	M 2.5 8 8 8 1.7 1.7 1.7 1.7	M 2.5 8 8 8 11.7 11.7 11.7 15.7 15.7 15.7 15.7 15.7	

Others

Part Name	Model Name
Finger lifter	AA-969
Lug back	AA-3310
Extension cable	AA-8801 (5m)
	AA-8802 (10m)
	AA-8803 (20m)
	AA-8804 (30m)
Gauge stand	ST-0230





ONO SOKKI

Outer appearance and specifications are subject to change without prior notice.
 URL: http://www.onosokki.co.jp/English/english.htm

U.S.A

Ono Sokki Technology Inc. 2171 Executive Drive, Suite 400 Addison, IL. 60101, U.S.A Phone: +1-630-627-9700

Phone: +1-630-627-9700 Fax: +1-630-627-0004 E-mail: info@onosokki.net http://www.onosokki.net

THAILAND

Ono Sokki (Thailand) Co., Ltd. 29/67 Moo 5 Tivanon Road, Pakkred, Nonthaburi 11120, Thailand

Phone: +66-2-964-3884 Fax: +66-2-964-3887 E-mail: osth_sales@onosokki.co.jp

P.R.CHINA

Ono Sokki Beijing Office Beijing Jing Guang Center 3510 Hu Jia Lou, Chao Yang Qu Beijing 100020, P.R.China Phone: +86-10-6597-3113

Fax : +86-10-6597-3113 Fax : +86-10-6597-3114 E-mail : onosokki@bbn.cn

WORLDWIDE

Ono Sokki Co., Ltd.

3-9-3 Shin-Yokohama, Kohoku-ku, Yokohama 222-8507, Japan Phone: +81-45-476-9712

Phone: +81-45-476-9712 Fax: +81-45-470-7244 E-mail: overseas@onosokki.co.jp

^{*2} When affixing a roller gauge head such as the AA-827/828 to a gauge with a measurement resolution of 1µm, there may be times when the precision specification cannot be achieved. The AA-827/828 uses a bearing, but as the gap cannot be eliminated, an error of approximately 10µm may appear.